

## SOUND LEVEL CALIBRATOR (1 kHz)

## NAGMAN 222

## SALIENT FEATURES

94dB and 114dB Sound Calibrator At 1KHz  
 Accurate And Simple to Use  
 Fits 1" & 1/2" Microphones Low Battery Indicator  
 Low Battery Indicator  
 IEC 60942(2003) Class 2, Ans1.40-1984

## GENERAL SPECIFICATIONS

<b>Output Sound Pressure Levels</b>	114dB and 94dB re 20uPa under reference Conditions
<b>Output Frequency</b>	1000Hz ±2%
<b>Total Harmonic Distortion</b>	< 4%
<b>Reference Condition</b>	Temperature 23°C(73°F)
<b>Static press</b>	103hpa Relative humidity 50%
<b>Accuracy of sound pressure level</b>	Understated reference environment conditions ±0.5Db
<b>Battery test</b>	Operate : Green LED indicator Low battery : Red LED indicator
<b>Battery</b>	9V battery
<b>Dimensions</b>	113x63x44 mm
<b>Weight</b>	Approx.170g



Typical Photo: Final Product look may vary

## STANDARD INCLUSIONS

- Basic Instrument with Battery
- Carrying Case
- Conformance Certificate
- Instruction Manual

## OPTIONAL

- Calibration Certificates are issued in Accordance with our Scope as granted by NABL per ISO/IEC 17025:2017 Standards

Manufactured by



NAGMAN INSTRUMENTS AND ELECTRONICS PRIVATE LIMITED

AN ISO 9001:2015 CERTIFIED COMPANY

170/1a2, Chennai Bangalore Highway Road, Sembarambakkam, Chennai – 600 123. INDIA

Phone - Domestic Sales 044-6677 7024, 044-6677 7005 Email: [mktgchennai@nagman.com](mailto:mktgchennai@nagman.com)

Phone - Export Sales +91-44-6677 7025. Email: [exports@nagman.com](mailto:exports@nagman.com)

[www.nagman.com](http://www.nagman.com)

Specifications subject to change owing to continuous development. Contact us for latest Datasheet.



NIE/TDS/222/01/26/00